

CERTIFICATE

We hereby certify for **BORO-8330™ (glass 8330) batch 8330S024** following values:

Hydrolytic resistance (glass grains test):

Acc. to	Value	Limit	Unit	Result
ISO 719	8	HGB 1 ≤ 31	µg Na ₂ O / 1 g glass	HGB 1
ISO 720	11	HGA 1 ≤ 62	µg Na ₂ O / 1 g glass	HGA 1
Ph. Eur. 3.2.1	0.018	Type I ≤ 0.1	ml 0.02 M HCl / 1 g glass	Type I
USP <660>	0.018	Type I ≤ 0.1	ml 0.02 M HCl / 1 g glass	Type I
JP 7.01	0.05	Containers to be fused ≤ 0.30	Containers not to be fused ≤ 2.00 ml 0.01 M H ₂ SO ₄ / 5 g glass	fulfilled

Arsenic extraction (glass tube):

Acc. to	Value	Limit glass containers	Unit
Ph. Eur. 3.2.1	< 0.01*	0.1	ppm As

* measured on a glass tube

All measurements according to the current edition / version of the respective reference.

SCHOTT AG
Business Unit Tubing
Site Mitterteich

Deputy Head of Analytical Services

Maximilian Zeitner
Maximilian Zeitner

SCHOTT AG
Business Unit Tubing
Site Mitterteich

Global Director Quality Management

Karsten Hennig
Dr. -Ing. Karsten Hennig

SCHOTT
glass made of ideas